
jorùs

CERTIFICATE OF ANALYSIS

Chromatogram



Quantitative Results

Detector A		
Compound Name	Concentration, %	
CBDV		
CBDA		
CBGA		
CBG		
CBD	3.333	
THCV		
CBN		
ТНС		
СВС		
THCA-A		
CBL		
CBDVA		
CBDB		
L		

C	-	·		+ : - · -
Sampl	e	into	rma	ition

Batch number:	BSA-36
Sample number:	M2059
Date of Analysis	2022 08 10
Date of Analysis	2022 08 10

Summary

Total CBD	3.33 %
Total CBD	33.3 <mark>3 mg/g</mark>

Instrumental and analytical conditions:

Sample preparation: 0.1 g of sample material was disolved in 10 mL of HPLC grade methanol. The solution was vortexed and centrifuged. Then the solution was dilluted to a final concentration. Quantification of cannabinoids was performed using standard calibration curve method. Equipment: Quantitative analysis was performed using Shimadzu Cannabis Analyzer for Potency an integrated HPLC system with built-in sample cooler, degasser, autoinjector and UV detector. NexLeaf CBX for Potency, 2.7 µm, 4.6 x 150 mm collumn coupled with NexLeaf CBXGuard collumn was eluted by using a mixture of mobile phase A (0.085 % phosphoric acid in water) and mobile phase B (0.085% phosphoric acid in Acetonitrile) with a flow rate of 1.6 mL/min at 35°C. Sample injection volume was set to 5 µL. Gradient program used - 70 % B for 3 min, 70-85 % B over 4 min, 85-95 % B over 0.01 min; 70% B for 1.99min. Data was analyzed using Shimadzu LabSolutions software.

UAB Naturus, Stasio Lozoraičio 31 A, LT-71286 Baltrušiai, Šakių raj., Lithu info@naturus.lt